

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 10 2009

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130129</u>	
Name <u>J Swartz</u> ID# <u>4314</u> Date <u>8/4/09</u>	
A	Agency <u>SCDL</u> Phone # <u>269-5592</u>
Instrument Location <u>SCDL</u>	
Alco S/N <u>SC07108</u> Target Value <u>.080</u> High Pressure <u>700 psi</u>	
B	Alco Test Values <u>.078</u> <u>.078</u> 1 st Alco 2 nd Alco
Signature <u>[Signature]</u>	
(OVER)	

bmb
9/10/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Oct, 2009.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130129

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130129

AUGUST 04, 2009

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: AAA1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .080

ALCO S/N: SCOTT08

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.06 in

ALCO TARGET	.080	13:21
BLANK TEST	.000	13:21
INTERNAL STANDARD	VERIFIED	13:22
ALCO TV 30.07 in	.078	13:22
BLANK TEST	.000	13:23
SUBJECT SAMPLE	.000	13:23
BLANK TEST	.000	13:24
ALCO TV 30.07 in	.078	13:24
BLANK TEST	.000	13:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
ATAMASTER cdm 130129

AUGUST 04, 2009

IME 13:18

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
EATERS	
AMPLE CHAMBER:	49c
REATH TUBE:	41c
AROMETER:	30.06 in
LOW DETECTOR:	OKAY
JMP	
IGH SPEED:	OKAY
ETECTOR:	OKAY
ILTERS:	OKAY
UARTZ STANDARD:	OKAY
ALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}